

# Optical Engineering

[SPIDigitalLibrary.org/oe](http://SPIDigitalLibrary.org/oe)

## **Toward superfast three-dimensional optical metrology with digital micromirror device platforms**

Tyler Bell  
Song Zhang

# Toward superfast three-dimensional optical metrology with digital micromirror device platforms

Tyler Bell and Song Zhang\*

Iowa State University, Department of Mechanical Engineering, Ames, Iowa 50011

**Abstract.** Decade-long research efforts toward superfast three-dimensional (3-D) shape measurement leveraging the digital micromirror device (DMD) platforms are summarized. Specifically, we will present the following technologies: (1) high-resolution real-time 3-D shape measurement technology that achieves 30 Hz simultaneous 3-D shape acquisition, reconstruction, and display with more than 300,000 points per frame; (2) superfast 3-D optical metrology technology that achieves 3-D measurement at a rate of tens of kilohertz utilizing the binary defocusing method we invented; and (3) the improvement of the binary defocusing technology for superfast and high-accuracy 3-D optical metrology using the DMD platforms. Both principles and experimental results are presented. © 2014 Society of Photo-Optical Instrumentation Engineers (SPIE) [DOI: [10.1117/1.OE.53.11.112206](https://doi.org/10.1117/1.OE.53.11.112206)]

**Keywords:** superfast three-dimensional optical metrology; binary defocusing; phase shifting; fringe analysis; structured light.

Paper 140086SSP received Jan. 17, 2014; revised manuscript received Feb. 26, 2014; accepted for publication Feb. 27, 2014; published online Apr. 1, 2014.

## 1 Introduction

Recent advancements in three-dimensional (3-D) computational methods have led to high-resolution real-time 3-D shape measurement techniques that have been successfully applied to numerous areas, including manufacturing, medical sciences, homeland security, and entertainment.

Three-dimensional shape measurement technologies have seen substantial growth over the past few decades, and numerous techniques have been developed to recover 3-D objects with different principles.<sup>1,2</sup> The most popular methods include the time of flight, laser triangulation, shape from focus and defocus, stereo vision, structured light, and digital fringe projection (DFP). Each technology was developed to solve the challenging problems of the time for some specific application. However, no existing technology can conquer all of those challenges in each domain, and thus selecting the proper 3-D shape measurement technique for a particular application is still necessary. The handbook<sup>3</sup> assembles a number of 3-D shape measurement techniques and puts them side by side, making it easier for users to determine the right method for their specific application.

The structured light method became one of the most important 3-D shape measurement technologies for both scientific research and industrial practices mainly because of its simplicity and speed.<sup>4</sup> Real-time high-speed 3-D shape measurement has become even more popular because the processing power of a regular, modern computer (even a tablet) can handle such large amounts of data.<sup>5</sup> In our view, “real-time 3-D shape measurement” includes three major components: acquisition, reconstruction, and redisplay all occur at speeds of 24 Hz or higher.


Even though there are real-time 3-D shape measurement techniques developed based on other principles, such as time of flight,<sup>6</sup> active stereo vision,<sup>7</sup> and structured light,<sup>8</sup> the DFP techniques stand out because of their overwhelmingly advantageous features (e.g., high spatial and temporal resolution)

over other types of techniques.<sup>1</sup> The DFP technique is special kind of structured light method in that the structured patterns vary sinusoidally and continuous in both  $u$  and  $v$  directions.

This article presents our decade-long research efforts toward superfast 3-D shape measurement using the digital micromirror device (DMD) platforms. First, we will present the first-ever high-resolution real-time 3-D optical metrology system using the DFP method that we developed: the system can simultaneously acquire, reconstruct, and redisplay 3-D shapes at 30 Hz with over 300,000 measurement points per frame.<sup>9</sup> We will then present our recent innovations on the digital binary phase-shifting technique<sup>10</sup> that has demonstrated its merits over the conventional sinusoidal phase-shifting method in terms of measurement speed and simplicity by achieving kilohertz-rate 3-D optical metrology without worrying about the nonlinearity of the projection system.<sup>11</sup> These speed breakthroughs were enabled by the digital light processing (DLP) Discovery platform and more recently the DLP LightCrafter and DLP LightCommander platforms. We have also found that the binary defocusing technique is not trouble free: the squared binary method has smaller measurement depth range than the conventional method. To address this challenge, we have developed methods to improve the measurement accuracy without sacrificing the measurement speed as well as methods to increase the measurement depth range without losing measurement quality. Specifically, this article will summarize the developments of the optimal pulse width modulation (OPWM)<sup>12</sup> and the binary dithering/halftoning techniques<sup>13</sup> for high-quality superfast 3-D shape measurement with the binary defocusing method. Principles of each technique will be presented, and experimental results will be shown to verify its performance.

It should be emphasized that the majority of the technologies covered in this article have already been developed and published in Chapter 9 of the Handbook;<sup>3</sup> this article serves as a sort of review, from our own experience and perspective, on how the real-time 3-D shape measurement technologies

\*Address all correspondence to: Song Zhang, E-mail: [song@iastate.edu](mailto:song@iastate.edu)



**Fig. 1** Example of the projected timing signal with different grayscale values. (a) Green = 255 (Red = Blue = 0); (b) Green = 128 (Red = Blue = 0); (c) Green = 64 (Red = Blue = 0); (d) Green = 0 (Red = Blue = 0).

evolve by leveraging the unique features of hardware technologies. We hope that this article provides the readers with a coherent piece of literature on the status of the most recent superfast 3-D shape measurement technologies, motivates them to further advance these technologies, and drives them to adapt the technologies for their specific applications.

Section 2 discusses the basics of the DLP technology and the phase-shifting algorithm we used. Section 3 presents the first real-time 3-D shape measurement system, which was developed by modifying the commercially available DLP projectors. Section 4 presents the recent superfast 3-D shape measurement development, which was accomplished by inventing the binary defocusing technique and leveraging recent DLP hardware technologies. Section 5 discusses some challenging problems to conquer to increase the value of the real-time 3-D shape measurement technologies in practical applications. Last, Sec. 6 summarizes this article.

## 2 Principle

### 2.1 Grayscale Image Generation Mechanism of DLP Technology

It is well known that the DLP projector produces grayscale values by time modulation.<sup>14</sup> Specifically, the DMD modulates the grayscale values by properly tilting the mirrors either into or away from the optical path at a rapid speed and with a proper timing ratio to produce the image's pixels. Figure 1 illustrates the waveform of the timing signal under different grayscale values. In these experiments, a photodiode sensor (Thorlabs FDS100, Newton, New Jersey) was used to detect the output light. The photoelectric current signal was then converted into a voltage signal and monitored by an oscilloscope. Only the green channel was used for these experiment to better visualize the effect. These experimental results demonstrated that when the projector is fed with a pure green image [RGB = (0, 255, 0)], the DMD is almost 100% ON. When the grayscale value was reduced to 128 and then 64, the irregular waveforms were generated with an approximate 50% and 25% ON time, respectively, during the projection cycle. Furthermore, if the grayscale value is 0, the DMD is 0% ON. These experimental results indicate that the full projection cycle is required in order to properly generate 8-bit sinusoidal patterns. However, if only 0 or 255 grayscale values are necessary, a partial projection cycle is sufficient.

### 2.2 Three-Step Phase-Shifting Algorithm

Phase-shifting algorithms have been extensively employed in optical metrology due to their speed, accuracy, and robustness to noise. Even though there are numerous phase-shifting algorithms that have been developed, a simple three-step phase-shifting algorithm is usually preferable for high-speed 3-D shape measurement, since this is the minimum number of patterns required to solve for the phase value pixel by pixel, albeit the random noise effect may be significant. A three-step phase-shifting algorithm with a phase shift of  $2\pi/3$  can be realized by capturing three fringe images with equal phase shifts, and these three images can be mathematically described as

$$I_1(x, y) = I'(x, y) + I''(x, y) \cos[\phi - 2\pi/3], \quad (1)$$

$$I_2(x, y) = I'(x, y) + I''(x, y) \cos[\phi], \quad (2)$$

$$I_3(x, y) = I'(x, y) + I''(x, y) \cos[\phi + 2\pi/3]. \quad (3)$$

In Eqs. (1)–(3),  $I'(x, y)$  is the average intensity,  $I''(x, y)$  the intensity modulation, and  $\phi(x, y)$  the phase to be solved for. The phase can be calculated by simultaneously solving these three equations,

$$\phi(x, y) = \tan^{-1} \frac{\sqrt{3}(I_1 - I_3)}{2I_2 - I_1 - I_3}. \quad (4)$$

Due to the nature of the arctangent function, the phase is wrapped with a range from  $-\pi$  to  $\pi$  with a modulus of  $2\pi$ . In order to obtain a continuous phase map, a spatial or temporal phase unwrapping algorithm is required. This phase unwrapping essentially detects and removes  $2\pi$  discontinuities by adding or subtracting integer multiples of  $2\pi$  (Ref. 15). Once the system is calibrated,  $(x, y, z)$  coordinates can be reconstructed from the phase.<sup>16</sup>

## 3 Real-Time 3-D Shape Measurement with a Modified DLP Projector

Our efforts toward superfast 3-D shape measurement started with modifying a DLP projector mainly because of the availability of hardware at the time and the limitations of the other hardware components. The commercially available DLP projectors typically generate color images that are fed to the projectors. The colors are generated by putting a rapidly

spinning color wheel into the optical path and properly synchronizing the color wheel with the computer-generated signal. The synchronization signal is usually picked up by a photosensor behind the color wheel with some associated electronics. Figure 2 shows photos of the projector we modified (Kodak DP900, Rochester, New York). The photosensor behind the color wheel is shown in Fig. 2(a); this is used to detect the color filters and therefore used for synchronization purpose. Figures 2(b) and 2(c), respectively, show the color wheel with and without color filters.

As discussed in Sec. 2.1, DLP works off the time modulation to generate grayscale images, and thus synchronizing the projector and the camera becomes extremely important to properly capture the desired fringe images used for high-quality 3-D shape measurement. Moreover, using color fringe patterns is not desirable for high-quality 3-D shape measurement,<sup>18</sup> so the elimination of color is vital. Furthermore, the commercially available DLP projectors are typically nonlinear devices (also called nonlinear gamma effect) to accommodate for human vision, but generating ideal sinusoidal patterns is required for high-quality 3-D shape measurement. Meanwhile, the unique projection mechanism of DLP technology, sequentially projecting red, green, and blue channels, permits the natural way of high-speed 3-D shape measurement if a three-step phase-shifting algorithm is used. In summary, to achieve high-speed 3-D shape measurement at a desired high quality with a commercially available DLP projector, the following hardware challenges need to be conquered: (1) elimination of the color wheel from the optical path; (2) precise synchronization between the camera and the projector; (3) corrections for the nonlinear gamma effects.


### 3.1 Real-Time 3-D Shape Acquisition

To conquer those aforementioned three challenges, modifying the commercially available digital video projector is required. The most important modifications we made were removing the color filters of the color wheel and supplying an external trigger signal such that the projector can still take the regular color image from the computer and project the red, green, and blue channels sequentially, albeit in monochromatic mode. The external trigger signal was generated with microcontroller to mimic the signal from the photodetector such that the projector cannot tell whether the signal was from the photodetector or our circuit and thus behaves independently of the signal source.

After these modifications, the projector projects a monochrome fringe image for each of the RGB channels sequentially. Each “frame” of the projected image is actually three separate images. By removing the color wheel and placing each fringe image in a separate channel, the projector can produce three fringe images at 80 Hz (240 individual color channel refresh rate). Therefore, if three fringe images are sufficient to recover one 3-D shape, the 3-D measurement speed is up to 80 Hz. However, due to the speed limit of the camera used and the asynchronous data transfer time requirement, it takes two projection cycles to capture three fringe images; thus, the measurement speed is 40 Hz (Ref. 9). Figure 3(a) shows the timing chart for the real-time 3-D shape measurement system.


It is important to note that, as indicated in Fig. 3(a), the timing of different color channels is different (red channel lasts the longest time, and blue channel lasts the shortest time) for the digital video projector we modified, making the precise synchronization between the projector and the camera more difficult because rapidly changing the camera exposure time from one frame to the next was not permitted. The synchronization between the projector and the camera requires a time-consuming trial-and-error process. It is also important to note that the projector also has a clear channel that is intended to increase the overall brightness of the projector, but this does not provide additional useful information for our 3-D shape measurement purpose.

As addressed in Sec. 2.2, three fringe images can be used to reconstruct one 3-D shape if a three-step phase-shifting algorithm is used. This perfectly fits into the DLP technology schema, where three patterns can be encoded into each of the three primary color channels of the projector. Since color fringe patterns are not desirable for 3-D shape measurement, we developed a real-time 3-D shape measurement system based on a single-chip DLP projector and white light technique.<sup>9</sup> Figure 4 shows the system layout. The computer-generated color-encoded fringe image is sent to a single-chip DLP projector that projects three color channels sequentially and repeatedly in grayscale onto the object. The camera, precisely synchronized with the projector, is used to capture three individual channels separately and quickly. By applying the three-step phase-shifting algorithm to three fringe images, the 3-D geometry can be recovered. Averaging three fringe images will result in a texture image that can be further mapped onto the recovered 3-D shape to enhance certain visual effects.



**Fig. 2** Color wheel and photosensor of the digital video projector. (a) The photosensor behind the color wheel; (b) color wheel of the projector; (c) color wheel after removing color filters. (Original figure was published in Ref. 17.)





**Fig. 3** Timing chart of the measurement system. (a) The first projector (Kodak DP900) was modified without firmware changes; (b) the second projector (PLUS U5-632h) was modified with an additional PLUS Vision firmware update.

### 3.2 Projector Nonlinear Gamma Correction

Since 3-D geometry is directly recovered from the phase, the phase quality is essential to the measurement accuracy: any noise or distortion on the phase will be propagated to the final 3-D measurement. Unlike a structured light system using a binary coding method, the DFP system requires calibrating for the nonlinearity of the projection and compensating for the errors associated with it. Even though nonlinearity correction is relatively mature with many methods developed,<sup>19–24</sup> our research found that the nonlinearity of the projector actually changes over time; this complicates the problem since a regular nonlinearity calibration is usually sufficient.<sup>5</sup> Nevertheless, good quality phase can be obtained after applying a proper nonlinear gamma correction algorithm.

### 3.3 Real-Time 3-D Data Processing and Visualization

As mentioned above, real-time 3-D shape measurement requires real-time 3-D shape acquisition, reconstruction, and redisplay. The processing time is tremendous for computing the wrapped phase point by point using a standard arctangent function, and when phase unwrapping is factored in as well, the challenge becomes even more difficult. These are just some of the reasons that real-time 3-D shape reconstruction was difficult about 10 years ago (around the year 2004). By developing a new method for computing wrapped phase,<sup>25</sup> and substantially optimizing our phase unwrapping algorithm, we successfully achieved real-time 3-D data processing for an image resolution of  $532 \times 500$  (Ref. 26). Figure 5 shows an example frame which is measuring human facial expressions in real time (40 Hz in this case).

There are, however, two issues associated with the first real-time 3-D shape measurement system we developed:<sup>26</sup>

1. *Inaccurate calibration method.* To achieve real-time 3-D data processing, the simple reference-plane-based calibration method was adopted. As a result, the recovered 3-D geometry could be significantly distorted if the object is away from the reference plane, or if the hardware system is not properly configured.<sup>16</sup>
2. *Low-phase quality.* The phase quality is reasonable but not the best in principle. This is because the three color channels have different illumination times, shown


in Fig. 3(a), but the camera uses the exposure time to capture these channels. The result is that not all channels are properly captured.

Further hardware modifications are required in order to solve for these two problems. Fortunately, PLUS (Tokyo, Japan) Vision agreed to do firmware modifications for one of their projector models, PLUS U5-632h. With the company's assistance, besides removing color filters, the modified projector only projects red, green, and blue channels without the clear channel, and all these channels have the same duration time. These modifications were highly desirable for high-quality 3-D shape measurement. Furthermore, because of these modifications, the projection speed can be at 120 Hz (or 360 Hz individual channel refreshing rate), speeding up the whole measurement (to 120 Hz). Figure 3(b) shows the timing chart for the second-generation real-time 3-D shape measurement system we developed.<sup>27</sup>

Besides improved phase quality, the new system also adopted the more complex, yet more accurate, calibration method we developed.<sup>16</sup> The new calibration method requires substantial computation power in order to achieve real-time 3-D data processing, making it almost impossible to achieve on a central-processing unit. Fortunately, graphics-processing unit (GPU) technologies were invented and quickly adopted.

By taking advantage of the processing power of the GPU, 3-D coordinate calculations can be performed in real time with an ordinary personal computer with an NVidia graphics card.<sup>28</sup> Moreover, because 3-D shape data is already on the graphics card, it can be rendered immediately without any delay. Therefore, by these means, real-time 3-D geometry visualization can also be realized in real time simultaneously. Also, because only the phase data, instead of the 3-D coordinates and the surface normals, are transmitted to the graphics card for visualization, this technique reduces the data transmission load on the graphics card significantly (approximately six times smaller). In short, by utilizing the processing power of GPU for 3-D coordinate calculations, real-time 3-D geometry reconstruction and visualization can be performed rapidly and in real time.

Figure 5 shows an experimental result of measuring a live human face. The right figure shows the real subject, and the left shows the 3-D geometry acquired and rendered on the computer screen at the same time. The simultaneous 3-D



**Fig. 4** Real-time three-dimensional (3-D) shape measurement system layout. (Original figure was published in Ref. 5.)

data acquisition, reconstruction, and display speed achieved is 30 frames/s with more than 300,000 points per frame being acquired.

These technologies, though successful, require substantial projector modifications, and sometimes these modifications are impossible without the projector manufacturer's involvement. Fortunately, with about a decade of effort on high-speed 3-D shape measurement, projector manufacturers are beginning to realize the opportunities in this field by producing affordable specialized projectors: LogicPD LightCommander being the first, and Texas Instruments (TI, Dallas, Texas) following up with its LightCrafter series.

#### 4 Superfast 3-D Shape Measurement with Binary Defocusing Techniques

As 3-D shape measurement technologies become more accurate and faster, such as with the real-time systems which have been proposed, the number of potential applications in the field continue to grow. These technologies have already been applied successfully to the medical field, the entertainment field, the manufacturing field, and many more. Up until now, the definition for a real-time system might be one that can capture at a rate of 30 or higher hertz. This speed enables a system to capture scenes or objects that are not too rapidly changing or moving around. A good example of this might be one's facial expression. If an object being captured does




**Fig. 5** Simultaneous 3-D data acquisition, reconstruction, and display at 30 fps. (Original figure was published in Ref. 5.)

move at a very fast pace, however, such as a speaking mouth or a live, beating heart, some interesting challenges need to be overcome to realize a kilohertz scanning rate. Fortunately, there have been advances in hardware platforms (e.g., DLP Discovery, DLP LightCommander, and the DLP LightCrafter); these advancements to the hardware allow for them to display 8-bit grayscale images at a few hundred hertz and allow for the display of binary images at an even faster kilohertz or higher rate. These advancements have driven us to invent the binary defocusing technique<sup>10</sup> such that a scanning at a kilohertz rate can be achieved with good quality.

Zhang et al.<sup>11</sup> have capitalized on the advancements made in the hardware platforms and successfully achieved a 3-D shape measurement speed of tens of kilohertz. Albeit fast, this new system comes with its own challenges and setbacks: (a) the presence of high-frequency harmonics impacts the measurement accuracy; (b) the generated binary pattern has an optimal quality at a certain range, and the region for this range is relatively small thus the result may be a less accurate depth measurement;<sup>29</sup> (c) conventional calibration methods take into account that the projector is in focus yet this new capture method relies upon the projector being defocused; the result is needing a more complex calibration technique;<sup>30</sup> and (d) high-quality fringe patterns at different spatial frequencies are hard to generate simultaneously.<sup>31</sup> To overcome some of these challenges, the 3-D shape measurement field has borrowed pulse width modulation (PWM) techniques from the Power Electronics field with Fujita et al.<sup>32</sup> and Yoshizawa and Fijita<sup>33</sup> implementing the PWM technique through hardware, Ayubi et al.<sup>34</sup> realizing the sinusoidal pulse width modulation technique, Wang and Zhang<sup>12</sup> realizing the OPWM technique, and Zuo et al.<sup>35</sup> further improving such a technique. These techniques each work to improve the phase quality, yet they too still face their own issues: especially if the stripes in the fringe images are either too narrow or too wide.<sup>36</sup>

Given the discrete nature of these fringe patterns, however, there might not be many improvements that can be made in this regard. It should be stated that these PWM developments are all one-dimensional. It would be logical then that certain limitations can be lifted if improvements can be made in two dimensions instead. This is the driving hypothesis behind Lohry and Zhang's<sup>37</sup> new technique to locally modulate the pixels as to emulate a triangular pattern. This is done with the aim to reduce the influence of high-frequency harmonics.

Other research has been done to improve measurement quality by adopting dithering techniques. Dithering techniques, also sometimes called half-toning, have been around since the 1960s (Ref. 38); they can be used to represent grayscale images with binary images and, given the fast speeds at which modern projectors can display binary images, this becomes important. In regard to dithering application in the 3-D shape measurement field,<sup>39</sup> Wang and Zhang's work<sup>13</sup> extended the simple Bayer-dithering technique and eventually the error-diffusion techniques.<sup>40-44</sup> It should be noted that these techniques improve the quality of measurement when the fringe stripes are wide; improvement is not as noticeable when the fringe stripes are narrow. It should also be described that the majority of these dithering techniques involve applying some sort of matrix or kernel to the generated grayscale images to result in a binary image.




**Fig. 6** Captured patterns with the fringe period  $T = 90$  pixels, when the projector is slightly defocused. (a) Squared binary pattern; (b) pulse width modulation pattern; (c) dithering pattern; (d) optimized-dithering pattern; (e)–(h) corresponding reconstructed 3-D results of (a)–(d). (Original figure was published in Ref. 49.)

The resulting image, however, may not be optimal as these “binarization” techniques may not take into account some of the inherent structures in the grayscale images being used (such as sinusoidal structures in fringe images). Optimizations can be had, then, if these unique sinusoidal structures within the fringe patterns are assumed and capitalized upon. Recently, we have developed several algorithms<sup>45–48</sup> to

noticeably improve the aforementioned approaches to dithering.

One example of our improvements can be seen in Fig. 6. A 3-D sculpture was measured with a fringe period of  $T = 90$  pixels and with the projector being slightly defocused. The first row of the results displays the captured structured images; the second row displays the rendered 3-D results.



**Fig. 7** Example of capturing live rabbit hearts with binary defocusing techniques. (a)–(d) Captured fringes of a live rabbit heart; (e)–(h) corresponding reconstructed 3-D results of (a)–(d). (Original figure was published in Ref. 49.)



The results show that when the projector is nearly in focus, the binary structure is clear for the squared binary patterns and that the optimized dithered patterns are sinusoidal in nature. It should also be noted that neither the squared binary method nor the PWM method generates high-quality 3-D results, whereas the dithering and optimized-dithering techniques perform well in the sculpture's reconstruction. The last result which should be derived from this experiment is that the optimized-dithering technique is visibly better than the regular dithering method.

A second example of our work in realizing a system which can perform kilohertz 3-D shape measurement can be seen in Fig. 7. This system captures live rabbit hearts as they beat at a rate of  $\sim 200$  beats/min (Ref. 39). This specific system is composed of the TI DLP LightCrafter projector and a Vision Research Phantom (Wayne, New Jersey) V9.1 high-speed CMOS camera. As in our aforementioned systems, the camera was triggered to capture by an external electronic circuit that senses the timing signal from the projector and synchronizes the system. The camera's resolution was  $576 \times 576$  and the projector's resolution was  $608 \times 648$ . The projector's rate at which it switched binary patterns to be displayed was 2000 Hz, and thus the camera was set to capture at 2000 Hz, as well. A two-wavelength phase-shifting technique was used with the shorter wavelength being the OPWN pattern<sup>12</sup> and longer wavelength being Stucki-dithered patterns.<sup>44</sup> The resulting images show the clear and dynamic motion of the rabbit heart as it was well captured. It was found that a speed of at least 800 Hz was required to capture the heart accurately without being distorted by motion, and thus without the employment of the abovementioned binary defocusing techniques this would not have been possible.

## 5 Discussion

The advances in real-time superfast 3-D shape measurement have been tremendous using the DLP technologies, yet, there are still some challenging problems which should be addressed. This section presents some of these challenges we have realized but have not been able to completely overcome.

### 5.1 Real-Time 3-D Shape Measurement with the Sinusoidal Fringe Projection Techniques

Even though the conventional methods of using sinusoidal fringe patterns have some challenges, their major advantage is still obvious; especially, when using the off-the-shelf projectors where the projector has a very large depth of focus range. Also, high-accuracy calibration can be achieved since projector-camera pair calibration techniques are rather mature. However, the major challenges are

- *Projection lens distortion.* Because of the limited fringe pattern used, the projection lens distortion is usually ignored when performing 3-D reconstruction. The projection lens distortion is usually not very severe, but if the application requires a high-accuracy measurement (better than submillimeter), such an issue should be taken into account.
- *Nonlinear response.* Though algorithms have been developed to compensate for the error caused by the nonlinear response of the system, the nonlinear gamma effect cannot be completely eliminated. It should be

noted that the nonlinear effect is caused by not only the projector's gamma but also the camera sensor.

- *Reliability.* The majority of commercially available projectors use plastic projection lenses with focal lengths that could change due to heating. This creates huge problems for high-accuracy 3-D shape measurements, since it is very difficult to precisely calibrate the projector's parameters at its measurement condition (e.g., temperature).
- *DMD transient effect.* Our recent research<sup>50</sup> found that the DMD does not respond instantaneously; this can potentially create problems for the sinusoidal method, where the DMD switches ON and OFF rapidly thus making it difficult to completely eliminate this problem.

### 5.2 Superfast 3-D Shape Measurement with Binary Defocusing Techniques

We have made some speed breakthroughs utilizing binary-defocusing techniques, yet a number of new problems arrive that affect the performance of the measurement system. The major issues associated with the superfast 3-D shape measurement techniques are

- *Out-of-focus projector calibration.* Unlike the conventional sinusoidal fringe projection method, the binary defocusing technique uses an out-of-focus projector to produce sinusoidal fringe patterns. State-of-the-art structured light system calibration<sup>16,51–57</sup> assumes that the projector is at least nearly focused; accurately calibrating the defocused projector remains an open problem.
- *Depth range.* The projector's optical engine is typically designed to generate large depth of focus, which is a great feature for conventional fringe projection methods where the projector is always nearly focused. Yet, the binary defocusing method requires the projector to be out of focus and thus the measurement range is substantially compromised. The question becomes: can we achieve the same depth range by minimally modifying the optics of the projection system?
- *High-order harmonics influence.* The binary defocusing technique essentially suppresses high-order harmonics with the defocusing effect, yet, it may not be absolute zero if the projector is nearly focused. We have developed a number of techniques to optimize the pattern such that any effects of high-order harmonics are minimized. However, we believe that there is a lot of room to further advance these methods.
- *Quantifying the level of defocusing.* The measurement quality of the binary defocusing method is highly dependent upon the defocusing level: too much or too little defocusing will result in a lower quality of measurement. The current practice heavily relies on experience; quantitatively measuring the level of defocusing such that the highest measurement quality can be achieved every time is still an open research problem for us.

### 5.3 Application Challenges

Real-time superfast 3-D shape measurement techniques cannot thrive and continue to mature without solving real-world



problems. There are some practical issues that challenge the majority of current real-time (even static) optical 3-D shape measurement techniques, however, including

- *Automatic exposure.* Auto-exposure is very common for two-dimensional (2-D) imaging techniques, so we assume this can be done easily for 3-D shape measurement techniques as well. However, as important as this feature is, given that a projector is involved and given that this projector is a digital device, some non-trivial issues arise.
- *Shiny part measurement.* Optical methods suffer if the sensor sees a saturated image or an image without enough intensity. Shiny parts create both problems due to very high surface contrast. We believe such a problem could be at least partially solved through the usage of high-dynamic range (HDR) methods as used in 2-D techniques, yet it remains an open problem on how to rapidly and automatically achieve HDR for 3-D techniques.
- *Harsh industrial environment usage.* One of the largest application areas for 3-D shape measurement is quality control in the manufacturing field. However, the real-world environment is not always pleasant: high temperatures, high vibrations, flash lighting, etc., and all these pose challenges for adopting 3-D optical shape measurement systems for application on-site.
- *Outdoor applications.* Yet another application of a superfast 3-D shape measurement system is to capture dynamics of real scenes; this might involve a system to capture both inside and outside of a room. Since this technique is optically based, sunlight could pose challenges.

## 6 Conclusions

This article has summarized the route we took to achieve real-time kilohertz 3-D shape measurement speeds by leveraging the DLP technologies that were invented by TI. These technologies have already seen extensive applications, and there are many more applications in which these technologies can be applied. The fundamental barriers of achieving superfast 3-D shape measurement were successfully overcome through different innovations, but there are still a number of challenges we are facing. We believe further innovations are required to allow real-time superfast 3-D shape measurement technologies to be adopted on a much larger scale.

## Acknowledgments

The work presented in this article was carried out by numerous collaborators, Professors Fu-Pen Chiang from Stony Brook University, Peisen Huang from Stony Brook University, and Shing-Tung Yau from Harvard University; as well as numerous students from our laboratory at Iowa State University, Laura Ekstrand, Shuanyan Lei, Beiwen Li, William Lohry, and Yajun Wang. The authors would like to thank all of them. This study was sponsored by the National Science Foundation (NSF) under grant numbers: CMMI-1150711 and CMMI-1300376. The views expressed in this chapter are those of the authors and not necessarily those of the NSF.

## References

1. J. Geng, "Structured-light 3D surface imaging: a tutorial," *Adv. Opt. Photonics* **3**(2), 128–160 (2011).
2. S. Gorthi and P. Rastogi, "Fringe projection techniques: whither we are?," *Opt. Laser Eng.* **48**(2), 133–140 (2010).
3. S. Zhang, Ed., *Handbook of 3D Machine Vision: Optical Metrology and Imaging*, 1st ed., CRC Press, New York (2013).
4. J. Salvi et al., "State of the art in structured light patterns for surface profilometry," *Pattern Recognit.* **43**(8), 2666–2680 (2010).
5. S. Zhang, "Recent progresses on real-time 3-D shape measurement using digital fringe projection techniques," *Opt. Laser Eng.* **48**(2), 149–158 (2010).
6. S. Hsu et al., "Performance of a time-of-flight range camera for intelligent vehicle safety applications," in *Advanced Microsystems for Automotive Applications 2006 VDI-Buch 2006*, pp. 205–219 (2006).
7. Z. Zhang, "Microsoft kinect sensor and its effect," *IEEE Multimedia* **19**(2), 4–10 (2012).
8. S. Rusinkiewicz, O. Hall-Holt, and M. Levoy, "Real-time 3D model acquisition," *ACM Trans. Graph.* **21**(3), 438–446 (2002).
9. S. Zhang and P. S. Huang, "High-resolution, real-time three-dimensional shape measurement," *Opt. Eng.* **45**(12), 123601 (2006).
10. S. Lei and S. Zhang, "Flexible 3-D shape measurement using projector defocusing," *Opt. Lett.* **34**(20), 3080–3082 (2009).
11. S. Zhang, D. van der Weide, and J. Oliver, "Superfast phase-shifting method for 3-D shape measurement," *Opt. Express* **18**(9), 9684–9689 (2010).
12. Y. Wang and S. Zhang, "Optimum pulse width modulation for sinusoidal fringe generation with projector defocusing," *Opt. Lett.* **35**(24), 4121–4123 (2010).
13. Y. Wang and S. Zhang, "Three-dimensional shape measurement with binary dithered patterns," *Appl. Opt.* **51**(27), 6631–6636 (2012).
14. L. J. Hornbeck, "Digital light processing for high-brightness, high-resolution applications," *Proc. SPIE* **3013**, 27–40 (1997).
15. D. C. Ghiglia and M. D. Pritt, Eds., *Two-Dimensional Phase Unwrapping: Theory, Algorithms, and Software*, John Wiley and Sons, New York (1998).
16. S. Zhang and P. S. Huang, "Novel method for structured light system calibration," *Opt. Eng.* **45**(8), 083601 (2006).
17. S. Zhang, "High-resolution, real-time 3D shape measurement," Ph.D. Thesis, Stony Brook University (2005).
18. J. Pan, P. S. Huang, and F.-P. Chiang, "Color phase-shifting technique for three-dimensional shape measurement," *Opt. Eng.* **45**(12), 013602 (2006).
19. S. Kakunai, T. Sakamoto, and K. Iwata, "Profile measurement taken with liquid-crystal grating," *Appl. Opt.* **38**(13), 2824–2828 (1999).
20. H. Guo, H. He, and M. Chen, "Gamma correction for digital fringe projection profilometry," *Appl. Opt.* **43**(14), 2906–2914 (2004).
21. B. Pan et al., "Phase error analysis and compensation for nonsinusoidal waveforms in phase-shifting digital fringe projection profilometry," *Opt. Lett.* **34**(4), 2906–2914 (2009).
22. P. S. Huang, C. Zhang, and F.-P. Chiang, "High-speed 3-D shape measurement based on digital fringe projection," *Opt. Eng.* **42**(1), 163–168 (2003).
23. S. Zhang and P. S. Huang, "Phase error compensation for a three-dimensional shape measurement system based on the phase shifting method," *Opt. Eng.* **46**(6), 063601 (2007).
24. S. Zhang and S.-T. Yau, "Generic nonsinusoidal phase error correction for three-dimensional shape measurement using a digital video projector," *Appl. Opt.* **46**(1), 36–43 (2007).
25. P. S. Huang and S. Zhang, "Fast three-step phase shifting algorithm," *Appl. Opt.* **45**(21), 5086–5091 (2006).
26. S. Zhang and P. Huang, "High-resolution, real-time 3-D shape acquisition," in *IEEE Comp. Vision Pattern Recognit Workshop*, Vol. 3, pp. 28–37, IEEE (2004).
27. S. Zhang and S.-T. Yau, "High-resolution, real-time 3-D absolute coordinate measurement based on a phase-shifting method," *Opt. Express* **14**(7), 2644–2649 (2006).
28. S. Zhang, D. Royer, and S.-T. Yau, "GPU-assisted high-resolution, real-time 3-D shape measurement," *Opt. Express* **14**(20), 9120–9129 (2006).
29. Y. Xu et al., "Phase error compensation for three-dimensional shape measurement with projector defocusing," *Appl. Opt.* **50**(17), 2572–2581 (2011).
30. L. Merner, Y. Wang, and S. Zhang, "Accurate calibration for 3d shape measurement system using a binary defocusing technique," *Opt. Laser Eng.* **51**(5), 514–519 (2013).
31. Y. Wang and S. Zhang, "Superfast multifrequency phase-shifting technique with optimal pulse width modulation," *Opt. Express* **19**(6), 5143–5148 (2011).
32. H. Fujita et al., "Three-dimensional profilometry using liquid crystal grating," *Proc. SPIE* **5058**, 51–60 (2003).
33. T. Yoshizawa and H. Fujita, "Liquid crystal grating for profilometry using structured light," *Proc. SPIE* **6000**, 60000H1 (2005).
34. G. A. Ayubi et al., "Pulse-width modulation in defocused 3-D fringe projection," *Opt. Lett.* **35**(21), 3682–3684 (2010).

35. C. Zuo et al., "Optimized pulse width modulation pattern strategy for three-dimensional profilometry with projector defocusing," *Appl. Opt.* **51**(19), 4477–4490 (2012).
36. Y. Wang and S. Zhang, "Comparison among square binary, sinusoidal pulse width modulation, and optimal pulse width modulation methods for three-dimensional shape measurement," *Appl. Opt.* **51**(7), 861–872 (2012).
37. W. Lohry and S. Zhang, "3D shape measurement with 2D area modulated binary patterns," *Opt. Laser Eng.* **50**(7), 917–921 (2012).
38. T. L. Schuchman, "Dither signals and their effect on quantization noise," *IEEE Trans. Commun. Technol.* **12**(4), 162–165 (1964).
39. Y. Wang et al., "3D absolute shape measurement of live rabbit hearts with a superfast two-frequency phase-shifting technique," *Opt. Express* **21**(5), 5822–5632 (2013).
40. W. Purgathofer, R. Tobler, and M. Geiler, "Forced random dithering: improved threshold matrices for ordered dithering," *IEEE Int. Conf. Image Process.* **2**, 1032–1035 (1994).
41. B. Bayer, "An optimum method for two-level rendition of continuous-tone pictures," in *IEEE Int. Conf. Commun.*, Vol. 1, pp. 11–15, IEEE (1973).
42. T. D. Kite, B. L. Evans, and A. C. Bovik, "Modeling and quality assessment of halftoning by error diffusion," *IEEE Int. Conf. Image Process.* **9**(5), 909–922 (2000).
43. R. Floyd and L. Steinberg, "An adaptive algorithm for spatial gray scale," *Proc. Soc. Inf. Disp.* **17**(2), 75–77 (1976).
44. P. Stucki, "Mecca: A multiple-error correcting computation algorithm for bilevel hardcopy reproduction," Technical Report, IBM Res. Lab., Zurich, Switzerland (1981).
45. W. Lohry and S. Zhang, "Genetic method to optimize binary dithering technique for high-quality fringe generation," *Opt. Lett.* **38**(4), 540–542 (2013).
46. J. Dai and S. Zhang, "Phase-optimized dithering technique for high-quality 3D shape measurement," *Opt. Laser Eng.* **51**(6), 790–795 (2013).
47. J. Dai, B. Li, and S. Zhang, "High-quality fringe patterns generation using binary pattern optimization through symmetry and periodicity," *Opt. Laser Eng.* **52**, 195–200 (2014).
48. J. Dai, B. Li, and S. Zhang, "Intensity-optimized dithering technique for high-quality 3D shape measurement," *Opt. Laser Eng.* **53**, 79–85 (2014).
49. B. Li et al., "Some recent advances on superfast 3D shape measurement with digital binary defocusing techniques," *Opt. Laser Eng.* **54**, 236–246 (2014).
50. Y. Wang et al., "Digital micromirror transient response influence on superfast 3D shape measurement," *Opt. Laser Eng.* **58**, 19–26 (2014).
51. Q. Hu et al., "Calibration of a 3-D shape measurement system," *Opt. Eng.* **42**(2), 487–493 (2003).
52. F. J. Cuevas, M. Servin, and R. Rodriguez-Vera, "Depth object recovery using radial basis functions," *Opt. Commun.* **163**(4), 270–277 (1999).
53. R. Legarda-Sáenz, T. Bothe, and W. P. Jüptner, "Accurate procedure for the calibration of a structured light system," *Opt. Eng.* **43**(2), 464–471 (2004).
54. P. Huang and X. Han, "On improving the accuracy of structured light systems," *Proc. SPIE* **6382**, 63820H (2006).
55. W. Gao, L. Wang, and Z. Hu, "Flexible method for structured light system calibration," *Opt. Eng.* **47**(8), 083602 (2008).
56. R. Yang, S. Cheng, and Y. Chen, "Flexible and accurate implementation of a binocular structured light system," *Opt. Lasers Eng.* **46**(5), 373–379 (2008).
57. Z. Li et al., "Accurate calibration method for a structured light system," *Opt. Eng.* **47**(5), 053604 (2008).

**Tyler Bell** is a human computer interaction graduate student in Dr. Song Zhang's 3-D Machine Vision Research Lab at Iowa State University. His background is in computer science, in which he earned his bachelor's degree, summa cum laude. Tyler's research interests include the miniaturization of real-time 3-D imaging systems, applying 3-D shape measurement systems to real-world problems, and developing interfaces for interacting with 3-D systems and their data.

**Song Zhang** is the William and Virginia Binger assistant professor of mechanical engineering at Iowa State University. He received his PhD degree from Stony Brook University in 2005. His major research interests include high-speed 3-D optical metrology and 3-D information processing. He has authored or co-authored over 100 research articles. He is the recipient of NSF CAREER award in 2012, and a fellow of SPIE.